

Object Color Estimation Using Digital Camera for Noncontact Imaging Applications (Case Study: Teeth Color Estimation)

Dr. Sinan Majid Abdul satar

Department of Laser and Optoelectronic Engineering, University of Technology/Baghdad.

Email:sinansma@yahoo.com

Dr. Mehdi Munshid shellal

Department of Laser and Optoelectronic Engineering, University of Technology/Baghdad.

Ahmed Chyad Kadhim

Department of Laser and Optoelectronic Engineering, University of Technology/Baghdad.

Zainab M. Abdul Ameer

Department of Conservative Dentistry, Collage of Dentistry, University of Baghdad/Baghdad.

Received on: 26/6/2012 & Accepted on: 10/1/2013

ABSTRACT

There has been a considerable effort in several fields, for objective color analysis and characterization. This paper study the problem of teeth color estimation using the most common color space in current use being the Commission International ed' Eclairage L*a*b* parameters of the captured images were derived from Photoshop software and saved as digital shade guides. A set of ceramic teeth shade was used, Images were captured with high resolution digital camera with LED flash and the data obtained was used to calculate the L*a* and b* values, these values conventionally obtained directly from RGB space. This paper proposed a method to estimating teeth color information from acquired camera data by calculating the L*a*b* values from the device independent color space the CIEXYZ space using transformation matrix which is involve calibration and characterization parts. The results showed significant correlation between data obtained from the digital camera using calibration and characterization process with that obtained from spectrophotometer, also the result showed that the digital camera can be used for color measurements in dental clinic, where the digital imaging could be an alternative for using colorimetric or spectrophotometric methods.

Keywords: Digital camera, digital shade guide, colorimeter, noncontact imaging, camera calibration.

تخمين لون الجسم باستخدام كاميرة رقمية للتطبيقات اللاتلامسية (دراسة الحالة: تخمين لون الاسنان)

الخلاصة

هناك جهد كبير في العديد من المجالات، بهدف تحليل وتوصيف اللون. هذا البحث يقدم دراسة لمشكلة تقدير لون الأسنان باستعمال نطاق اللون الأكثر شيوعاً في الاستخدام الحالي وهو متغيرات الـ $L^*a^*b^*$ للصور التي تم التقاطها يتم اشتقاقها من برنامج فوتوشوب وحفظها كمرشد رقمي. لقد تم استخدام مرشد الأسنان السيراميكي، والتقاط الصور بكاميرا رقمية عالية الدقة مع فلاش LED وقد تم استخدام البيانات التي تم الحصول عليها لحساب قيم الـ $L^*a^*b^*$ ، هذه القيم يتم الحصول عليها عادةً مباشرة من الفضاء RGB. لقد اقترح هذا البحث طريقة لتخمين المعلومات اللونية للأسنان من بيانات الكاميرا بواسطة حساب قيم الـ $L^*a^*b^*$ من الفضاء CIE XYZ باستخدام مصفوفة التحويل التي تتضمن على معايرة وتوصيف. أظهرت النتائج ارتباط كبير بين البيانات التي تم الحصول عليها من الكاميرا الرقمية باستخدام المعايرة وعملية التوصيف مع التي تم الحصول عليها من Spectrophotometer وكذلك أظهرت النتائج إمكانية استخدام الكاميرا الرقمية للقياس اللون في عيادة لطب الأسنان، حيث التصوير الرقمي يمكن أن يكون بديلاً عن استخدام الأساليب اللونية أو طيفية.

INTRODUCTION

One of many challenges in cosmetic dentistry is to accomplish appropriate and satisfactory reproduction of natural shade of teeth and in that way to make successful restoration [1]. Continuous technological improvements of dental materials and consequent aesthetic enhancement of direct and indirect restorations require trained practitioner able to choose the right shade. Achieving the satisfying morphological, optical and biological form of a restoration is one of the most important goals of esthetic dentistry and dentistry in general [2].

Instrumental assessment of color is considered to be better than subjective visual color matching methods [3]. However, the high cost hinders their common use. In addition to expensive colorimeters, digital images captured with a digital camera and subsequently analyzed using photo editing software has garnered more attention for color assessment [4-8].

Several studies have shown the potential of digital cameras for dental color matching [5-12]. The scene captured by an image sensor of a digital camera can be decomposed into red, green, and blue (RGB) components and translated into digital information. However, the color information received from a digital camera is device dependent. Images produced via a digital camera are usually affected by the light source, reflection spectra of the objects, the photo sensor of the digital camera, and how the image is processed and rendered by the digital camera.

In general, these captured signals are much influenced by the illumination under which the image is taken [8]. To compensate for color differences caused by the colorcast

of various light sources, an automatic white balance (AWB) mechanism is usually employed in most high-end digital cameras. Similar to the human eye, digital cameras provide an AWB function to adjust the color of pixels under different lighting conditions. The most widely used algorithm for AWB is based on the gray world assumption, which seeks to equalize the mean of RGB channels. Another commonly used algorithm incorporates the white world assumption, which states that the RGB values of the brightest point in the image should be the same [13]. However, AWB often fails in situations where the light source has a particular cast or if the captured image does not contain natural white [14]. This is especially important in dental color matching, since various illuminants are used in dental clinics, and the captured images are usually focused in small areas composed largely of a tooth's structure.

To make a successful restoration, the purpose of this study is to find a non contact method to calculate teeth color using digital camera.

THEORETICAL BACKGROUND

Whenever a picture for tooth samples is captured, a light is incident on the tooth surface this incident light is reflected by the tooth onto the camera sensor as a function of the reflectance spectrum of the target then the R, G, B values measured by the camera. This model is illustrated in the Figure (1).




Figure (1) Light source – Reflectance - Sensor model.

As shown [L] is a diagonal matrix represents the normalized light source spectrum values. The reflectance matrix [R] is a matrix with the normalized reflectance spectra on the camera sensor, and the [S] sensor sensitivity function matrix of the camera sensor. If [K] is the response matrix which contains the RGB values, it is given by:

$$[K]^T = [R] [L] [S] \quad \dots\dots\dots (1)$$

Where

$$K^T = \begin{bmatrix} R_1 & G_1 & B_1 \\ \vdots & \vdots & \vdots \\ R_n & G_n & B_n \end{bmatrix} \quad \dots\dots\dots (2)$$

$$R = \begin{bmatrix} r_1(\lambda_1) & \dots\dots\dots r_1(\lambda_m) \\ r_2(\lambda_1) & \dots\dots\dots r_2(\lambda_m) \\ \vdots & \vdots \\ r_n(\lambda_1) & \dots\dots\dots r_n(\lambda_m) \end{bmatrix} \quad \dots\dots\dots (3)$$

$$L = \begin{bmatrix} l(\lambda_1) & \dots\dots\dots \\ \dots\dots\dots l(\lambda_2) & \dots\dots\dots \\ \vdots & \vdots \\ \dots\dots\dots l(\lambda_m) \end{bmatrix} \quad \dots\dots\dots (4)$$

$$S = \begin{bmatrix} s_1(\lambda_1) & s_2(\lambda_1) & s_3(\lambda_1) \\ s_1(\lambda_2) & s_2(\lambda_2) & s_3(\lambda_2) \\ \vdots & \vdots & \vdots \\ s_1(\lambda_m) & s_2(\lambda_m) & s_3(\lambda_m) \end{bmatrix} \quad \dots\dots\dots (5)$$

n represents the number of color patches on the target. s_1, s_2, s_3 are the spectral sensitivity functions, r_1 to r_n are the reflectance spectra of n patches and l is the light source spectrum values; defined for each wavelength λ .

FINDING THE CALIBRATION MATRIX T

Matrix [K] contains the RGB values of the camera as predicted by the model just described. However these RGB values belong to a device dependent color space. In order to determine the actual colors that the camera RGB values represent, it is necessary to find a transformation that transforms them to device independent space coordinates as shown in Figure (2):




Figure (2) the transformation matrix.

The device independent color space in this case is the CIE XYZ space and the transformation matrix will be called $[T]_{\text{model}}$. These XYZ coordinates corresponding to the color patches are stored in matrix $[F]$. To determine the transformation from camera RGB to XYZ, the following set of equations in an over determined system was used:

$$[F]^T = [T]_{\text{model}} [K] \quad \dots\dots\dots (6)$$

Where,

$$F = \begin{bmatrix} X_1 & X_2 & \dots & X_n \\ Y_1 & Y_2 & \dots & Y_n \\ Z_1 & Z_2 & \dots & Z_n \end{bmatrix} \quad \dots\dots\dots (7)$$

$$T = \begin{bmatrix} m_{11} & m_{12} & m_{13} \\ m_{21} & m_{22} & m_{23} \\ m_{31} & m_{32} & m_{33} \end{bmatrix} \quad \dots\dots\dots (8)$$

The transformation matrix $[T]_{\text{model}}$ could be estimated by using the least squares approximation as following:

$$[T]_{\text{model}} = [F][K]^T ([K]^* [K]^T)^{-1} \quad \dots\dots\dots (9)$$

After that conversion from RGB to XYZ is performed, the $L^* a^* b^*$ tri-stimulus values from the XYZ is determined using the standard transformation equations [15].

The $L^* a^* b^*$ values have a perceptual meaning: L^* is the lightness which relates to the physical intensity of color, while a^* and b^* are coordinates on the red-green and yellow-blue axis respectively. This scheme is designed such that a constant difference in color, ΔE , is defined by Euclidean distance in the space, thus:

$$\Delta E = \sqrt{\Delta L^{*2} + \Delta a^{*2} + \Delta b^{*2}} \quad \dots\dots\dots (10)$$

As given in [1] the transformation can be divided into two parts:

1. Calibration
2. Characterization

Calibration

This process involves determining the nonlinear relation between the camera RGB values measured after gamma and shading correction in the camera and the linear RGB values which represent the response of the camera sensor system. In addition the calibration includes the process of estimation and correction of distortions caused by a

camera lens Camera calibration, in general, can involve estimating two types of parameters: Internal and external camera parameters. The internal parameters determine the relation between coordinates of points in the scene and their coordinates in the image, while the external parameters characterize the relations between the camera and the scene. Distortion is the aberration that is suffered by the ray that goes through the center of the lens [16]. Mainly two distortions can be considered: radial and tangential distortion.

Tangential distortion arises of misalignment of lens centers in a system of lenses. Moreover Tsai showed [17] that it is the radial distortion that is the most critical non-linear effect of the lens while calibrating a camera. Because of various constraints in the lens manufacturing process, straight lines in the world imaged through real lenses generally become somewhat curved in the image plane. The radial distortion that causes the image to bulge toward the center is called the barrel distortion, and distortion that causes the image to shrink toward the center is called the pincushion distortion. To do this calibration and removes these distortions a matlab functions are written for this purpose.

Characterization

Characterization is the process of finding the transformation from device dependent color space to device independent color space. They are of two types namely, forward characterization and inverse characterization.

MATERIALS AND METHODS

A set of ceramic teeth shade of A1-A3, A3.5, A4, B1-B4, C1-C4, D2, and D3 was used as shown in Figure (3). Images were captured with high resolution digital camera of five Megapixel and Carl Zeiss Optics with dual LED flash.



Figure (3) Ceramic teeth shade.


Six teeth samples were placed on black cloth backing when the digital images were taken as shown in Figure (4). The digital images were taken and saved in TIF format. The resolution used was (2592×1944) pixels.



Sample 1



Sample 2

**Figure (4) Teeth samples.**

Six different independent images were captured and retrieved on a 24-bit resolution screen then analyzed using Adobe Photoshop CS5.1 and Matlab 8. The color space of the software was set the same as the default color space of the camera. First the image mode was changed from RGB to $L^*a^*b^*$ (Lab_{dir}). A standardized square area of 36 mm^2 in area at the center of each sample image was cropped for analysis. The $L^*a^*b^*$ values of these area were measured six times for each samples, and the mean values were recorded.

RESULTS

The standard $L^*a^*b^*$ values of ceramic teeth shade measured by spectrophotometer as shown in table1. The $L^*a^*b^*$ color values first obtained directly from the RGB (Lab_{dir}) from the photoshop program for each sample, then these $L^*a^*b^*$ color values ($Lab_{Cal.}$) obtained from XYZ after calibration and characterization process.

**Table (1) The standard $L^*a^*b^*$ values of ceramic teeth shade
Measured by spectrophotometer.**

Ceramic shade	L^* value	a^* value	b^* value
A1	76.25	0.40	10.00
A2	73.10	0.65	12.75
A3	72.50	0.65	13.17
A3.5	69.50	2.80	18.00
A4	67.50	2.00	16.50
B1	76.25	-0.85	8.00
B2	75.00	-0.85	11.00
B3	71.25	0.00	14.80
B4	71.80	1.00	18.50
C1	75.60	-0.40	10.00
C2	70.50	0.85	12.75
C3	70.65	1.00	13.00


C4	66.75	1.30	12.96
D2	73.75	0.00	8.00
D3	70.60	0.65	11.00

Mean values of $L^*a^*b^*$ of ceramic shade is plotted in Figure (5). Lab_{dir} and Lab_{Cal} derived from digital images were compared. The correlation coefficients as shown in Table (2) below:

Table(2) The correlation between the standard $L^*a^*b^*$ values of ceramic teeth shade and that using direct method and the proposed after calibration and characterization process.

Parameter	Correlation Lab_{dir}	Correlation Lab_{Cal}
L^* _value	0.6537	0.9487
a^* _value	0.8986	0.9234
b^* _value	0.9152	0.9213

The proposed method of obtain digital shade guides using Lab_{Cal} had a greatest mean match than that using Lab_{dir} it is clear that it had significantly high correlations to the standard $L^*a^*b^*$.






Figure (5) Mean values of $L^*a^*b^*$ of ceramic shade.

CONCLUSIONS

A new method is proposed for finding teeth color using digital camera instead of conventional human visual methods using physical shade guide. The Lab_{cal} values of tooth image samples shows significantly high correlations to the standard $L^*a^*b^*$ values. The digital camera can be used for color measurements in the dental clinic.

REFERENCES

- [1].Aleksandra A., Dubravka M., Branislava P., Milica ., Valentina V., Tanja P.. Application of Advanced Technologies in Shade Matching – A Spectrophotometer. Serbian Dent.J . 2010; 57:225-227.
- [2]. Bayindir F, Kuo S, Johnston W, Wee A. Coverage error of three conceptually different shade guide systems to vital unrestored dentition.J Prosthet Dent. 2007; 98:175-85.

- [3]. Paul S, Peter A, Pietrobon N, Hämmerle CH Visual and spectrophotometric shade analysis of human teeth. *J Dent Res* 2002; 81:578–582
- [4]. Cal E, Sonugelen M, Güneri P, Kesercioglu A, Köse T. Application of a digital technique in evaluating the reliability of shade guides. *J Oral Rehabil* 2004; 31:483–491.
- [5]. Jarad FD, Russell MD, Moss BW . The use of digital imaging for colour matching and communication in restorative dentistry. *Br Dent J* 2005; 199:43–49.
- [6]. Wee AG, Lindsey DT, Kuo S, Johnston WM . Color accuracy of commercial digital cameras for use in dentistry. *Dent Mater* 2006; 22:553–559.
- [7]. Smith RN, Rawlinson A, Lath DL, Brook AH. A digital SLR or intra-oral camera: preference for acquisition within an image analysis system for measurement of disclosed dental plaque area within clinical trials. *J Periodontal Res* 2006; 41:55–61
- [8]. Oi-Hong T, Yu-Lin L, Yi-Ching H, I-Chiang C, Shyh-Yuan I. Development of digital shade guides for color assessment using a digital camera with ring flashes. *Clin Oral Invest* 2011; 15:49–56.
- [9]. Chu SJ, Tarnow DP. Digital shade analysis and verification: a case report and discussion. *Pract Proced Aesthet Dent* 2001; 13:129–136, quiz 138.
- [10]. Dozić A, Kleverlaan CJ, Aartman IH, Feilzer AJ Relation in color of three regions of vital human incisors. *Dent Mater* 2004;20:832–838.
- [11]. Smith RN, Collins LZ, Naeeni M, Joiner A, Philpotts CJ, Hopkinson I, Jones C, Lath DL, Coxon T, Hibbard J, Brook AH . The in vitro and in vivo validation of a mobile non-contact camera-based digital imaging system for tooth colour measurement. *J Dent* 2008;36:(Suppl 1) 15–20.
- [13]. Ebner M, Tischler G, Albert J. Integrating color constancy into JPEG2000. *IEEE Trans Image Process* 2007; 16:2697–2706.
- [14]. Lukac R . New framework for automatic white balancing of digital camera images. *Signal Process* 2008; 88:582–593.
- [15] Hunt RWG. “ Measuring color” 3rd ed. Fountain press. 1988.
- [16] MATTIAS MARDER. Comparison of Calibration Algorithms for a. Low-Resolution, Wide Angle, 3D Camera. MATTIAS MARDER. Master of Science Thesis. Stockholm, Sweden 2005-03-18.
- [17] R. Y. Tsai. A versatile camera calibration technique for high accuracy 3d machine vision metrology using off-the-shelf tv cameras and lenses. *IEEE J. Robotics and Automation*, 3(4):323.344, 1987.